









CONSTGLASS






Table of results



1-Pilot object

Pilot object:	Cologne Cathedral s XXII, south side aisle, "Adoration of the Magi" (1846)		
Picture	 <p>internal face</p>  <p>internal face</p>  <p>external face</p>	<p>Identification of the panel: s XXII (south side aisle) internal face</p> <p>Treatment and other materials: - the last conservation campaign of 1981/82 was carried out by an outside company.</p> <p><u>Product:</u> - BS 31 is a methyl silicone resin (50% Toluene + 2% hardener TU 2) (producer: Wacker Chemie, München, Germany); - SH 1 is a thermoplastic black coloured two component epoxy resin (producer: Glasbau Hahn, Frankfurt).</p> <p><u>Application:</u> The resin was applied with a brush after bonding of cracks and retouching. The application was carried out either partial or all over the surface. During the application ethyl acetate was used to prolong the setting time. - the samples are completely coated on the internal face; the external face was only coated along and over the cracks. - the sample has been divided in 3 separate pieces after the cracks had been opened. - the cracks were bonded with SH 1. - the panel was framed with U-profiles of brass, sealed with an industrial silicone for sealants. - on the linear edge of the sample still sticks an old adhesive tape.</p>	 <p>sampling area</p>  <p>internal face sample - Col_BS31_1 sample - Col_BS31_2 sample - Col_BS31_3</p>  <p>external face sample - Col_BS31_1 sample - Col_BS31_2 sample - Col_BS31_3</p>

	<div>  </div> <div> CONSTGLASS </div>	
	<div>Table of results</div>	

Remarks	<div> <p>All in all, the applied BS 31 seems to be in a rather sufficient condition. Only in some areas where the material had been thicker applied, the BS 31 has a milky aspect, looks crizzled and starts to delaminate. Obviously this delamination does not damage the original glass surface and the original paint, only the retouches – made immediately before the coating with BS 31 – supposedly will come off with the BS 31. Nevertheless, from the conservator's point of view there is no necessity to remove the BS 31 at the moment.</p> </div>
----------------	--



CONSTGLASS




Table of results



2-Results (first sample)

sample reference: Col_BS31_1 (Elettra)
light coloured glass with BS 31 on the internal face

Questions	Techniques	Answers
<p>Morphology</p> <ul style="list-style-type: none"> - What about the today's condition of BS 31? - What about the adhesion between glass and BS 31? - Did the BS 31 penetrated the paint-layer, is there a reaction between BS 31 and the paint layer? - What about the condition of the other materials?  <p>Col_BS31_1 (internal face)</p>	<p>Optical Microscope (DBV)</p>	<p>The internal face coated with BS 31 appears in a sufficiently good condition. Only some parts of the coating (thickly applied) have a milky aspect as if air bubbles were enclosed. Under the microscope crizzled areas are visible.</p> <p>The crack bonding is stable. To separate the sample, cracks had to be opened.</p> <p>The silicone has a good adhesion to the surface (in- and external face).</p> <p>The old cello-tape was used to fix the instable glass-edges during the last restoration. The tape is now brittle and lost adhesion.</p>



CONSTGLASS



Table of results



Typical phenomenon of the BS 31 treatment, exemplified by another area of the treated panel



internal face



internal face



external face

Where BS 31 delaminates, the areas have a milky aspect. This occurs mainly in the thicker applied parts of the coating and close to the cracks. In some areas the BS 31 coating has cracks.

SEM
(Institute)

n/a

**Desktop
tomography**
(Institute)

n/a (only Col_BS31_2)




CONSTGLASS




Table of results



Phase-contrast tomography on Synchrotron (Elettra)



Phase-contrast tomography: The compact glass with traces of metal, the silicone and the tape is very good visible.



3D-rendering of the same sample as above.

Chemical Composition
Is there any alteration of BS 31?

SEM/EDX
(Institute)

for organic materials this analytic device is not suitable.

Organic component composition

FTIR
(Institute)

for questions regarding the alteration of the product no reference spectrum of unaltered material for comparison issues is available, the product is not available any more.

RAMAN
(Institute)

for questions regarding the alteration of the product no reference spectrum of unaltered material for comparison issues is available, the product is not available any more.







CONSTGLASS



Table of results









Microbiology	Molecular biology ATP measurements (Institute)	not foreseen
Reversibility <i>-How can the film be removed without damage for glass and paint-layer?</i> <i>-Which method and what solvent can be used?</i>	Test studies elimination (DBV) The reversibility test was carried out <u>not</u> on the separated and scientifically investigated sample but on another comparable segment in the coated panel.	Tests with various solvents and solvent mixtures were carried out. Best result: By applying a compress with ethanol (gel with 5% Klucel® M), for 2x2 minutes and a subsequent cleaning with cotton swabs the BS 31 coating could visibly be removed.
 external face		
Re-treatability	Test studies re-treatability	not intended

	 CONSTGLASS	
	Table of results	

2-Results (second sample)

sample reference:	Col_BS31_2 (UGhent) light coloured glass with BS 31 on the internal face
--------------------------	--

Questions	Techniques	Answers
Morphology - What about the today's condition of BS 31? - What about the adhesion between glass and BS 31? - Did the BS 31 penetrated the paint-layer, is there a reaction between BS 31 and the paint layer? - What about the condition of the other materials? 	Optical Microscope (DBV)	The internal face coated with BS 31 appears in a sufficiently good condition. Only some parts of the coating (thickly applied) have a milky aspect as if air bubbles were enclosed. Under the microscope crizzled areas are visible. The crack bonding is stable. To separate the sample, cracks had to be opened. The silicone has a good adhesion to the surface (in- and external face). The old cello-tape was used to fix the instable glass-edges during the last restoration. The tape is now brittle and lost adhesion.
Col_BS31_2 (internal face)	SEM (Institute)	not foreseen
	Desktop tomography (UGhent)	  3D rendered volume of the glass fragment. The silicone is rendered orange, the glass green and the paint layer red.    blue: glass yellow: silicone (sealant) red line: paint layer brown: cello tape



CONSTGLASS



Table of results



	Phase-contrast tomography on Synchrotron (Institute)	n/a (only Col_BS31_1)
Chemical Composition <i>Is there any alteration of BS 31?</i>	SEM/EDX (Institute)	for organic materials this analytic device is not suitable.
Organic component composition	FTIR (Institute)	for questions regarding the alteration of the product no reference spectrum of unaltered material for comparison issues is available, the product is not available any more.
	RAMAN (Institute)	for questions regarding the alteration of the product no reference spectrum of unaltered material for comparison issues is available, the product is not available any more.
Microbiology	Molecular biology ATP measurements (Institute)	not foreseen
Reversibility	Test studies elimination	see above: Col_BS31_1 (first sample)
Re-treatability	Test studies re-treatability	not intended



CONSTGLASS



Table of results



Organic component composition	FTIR (Institute)	for questions regarding the alteration of the product no reference spectrum of unaltered material for comparison issues is available, the product is not available any more.
	RAMAN (Institute)	for questions regarding the alteration of the product no reference spectrum of unaltered material for comparison issues is available, the product is not available any more.
Microbiology	Molecular biology ATP measurements (Institute)	not intended
Reversibility	Test studies elimination	see above: Col_BS31_1 (first sample)
Re-treatability	Test studies re-treatability	not intended